
Report on Comparison Measurement

Phosphate buffer
2010 / 2011

1 Objectives

The comparison measurement was carried out in preparation of application for accreditation / CMC entries for the pH calibration laboratory of SNM INDECOPI.

2 Participants

Pilot laboratory

DKD-K-06901

ZMK -ANALYTIK- GmbH

Ortsteil Wolfen

P-D ChemiePark Bitterfeld-Wolfen

Areal A, Filmstr. 7

06766 Bitterfeld-Wolfen

Contact person: *Diana Jehnert (Deputy Head of DKD-K-06901)*

e-mail: info@zmk-wolfen.de

Participant

Servicio Nacional de Metrología – INDECOPI

Especialista 2

Lima, Peru

Contact person: *Galia Ticona*

e-mail: gticona@indecopi.gob.pe

3 Time schedule

Start of comparison measurement: October 2010

Submission of results by participant: March 2011

Evaluation of results: April 2011

4 Calibration object

Measuring object:	pH buffer solution / Phosphate
Nominal value:	pH 7
Prepared by:	ZMK -ANALYTIK- GmbH
Lot No.:	F1589786936 // A0135473017
Bottle No. (Participant)	13 and 14
Date of bottling:	15 th October 2010
Date of calibration (by ZMK):	18 th October 2010
Measuring temperature:	25°C

Homogeneity of the batch was determined before the comparison measurement was started. SNM INDECOPI received 2 bottles of 250 ml (bottle no. 13 and 14).

5 Applied calibration methods

5.1 DKD-K-06901 / ZMK -ANALYTIK- GmbH

For the determination of the reference value the **Differential potentiometry** was used by the pilot laboratory. The measuring cell is constructed according to Baucke [1]. The electrode system consists of two platinum electrodes that are surrounded by hydrogen. The measuring cell is placed in a thermostatic bath with known stability and homogeneity. A certified reference buffer solution was used as standard (Lot. No. 81680/73050, Batch No. 32, $U(k=2)=0.003$).




Fig. 1 Baucke cell design

[1] F.G.K. Baucke, J. Electroanal. Chem. 368 (1994), 67-75

5.2 SNM INDECOPI

The participant took part in the comparison measurement with two calibration methods.

Differential potentiometry

The measuring cell is constructed according to Baucke [1].

The electrode system consists of two platinum electrodes that are surrounded by hydrogen.

The measuring cell is placed in a thermostatic bath with known stability and homogeneity.

The potential difference was measured with a precision multimeter (Agilent 34420A).



Fig. 2 Baucke cell of SNM INDECOPI

A certified reference material with an uncertainty of $U(k=2)=0.003$ was used as standard:

Phosphate Lot. No. 73050/81680 certificate No.: 10-0241/DKD-K-06901/09-02

Multipoint calibration by glass electrode system

The calibration was carried out using a glass electrode measuring system.

Glass electrode:	Type:	Inlab Science
	Manufacturer:	Mettler Toledo
pH Meter:	Type:	PP50
	Manufacturer:	Sartorius
	Resolution:	0.1 mV

The following five certified reference materials were used as standards:

Tetraoxalate	Lot. No. 41908152	certificate No.: 10-0239/DKD-K-06901/09-02
Phthalate	Lot. No. A915674743	certificate No.: 10-0240/DKD-K-06901/09-02
Phosphate	Lot. No. 73050/81680	certificate No.: 10-0241/DKD-K-06901/09-02
Borate	Lot. No. 71840	certificate No.: 10-0242/DKD-K-06901/09-02
Carbonate	Lot. No. DMR-324-Ia / DMR-324-IIa	certificate No.: CNM-MR-630-0459/2009

The EMF of each reference solution and the sample was measured.

Using the calibration results a regression line and its parameters were determined.

These parameters were used to calculate the pH value of the unknown sample.

The calibration was carried out in a water bath with known stability and homogeneity.

6 Measuring results

The calibration results were reported in evaluation excel-sheets stating all raw data, calculation results and measuring uncertainties. The results of the comparison measurement are summarized in table 1.

In order to compare the results the value of E_n was calculated (see equation 1).

$$E_n = \frac{|x_{lab} - x_{ref}|}{\sqrt{U(x_{lab})^2 + U(x_{ref})^2}} \quad (1)$$

This value represents the deviation between the measuring results x_{lab} and x_{ref} of the participating laboratory and the pilot laboratory under consideration of the measuring uncertainty. For acceptable measurements the value of E_n must be less than 1.

Table 1: comparison measurement results

Measuring temperature	ZMK (DKD-K-06901)			SNM INDECOPI			E_n
	Calibration method	Reference pH value x_{ref}	exp. uncertainty U_{ref}	Calibration method	pH value x_{lab}	exp. uncertainty U_{lab}	
25 °C	Differential potentiometry	6.866	0.003	Differential potentiometry	6.865	0.003	0.24
25°C	Differential potentiometry	6,866	0.003	Multipoint calibration	6,87	0.02	0.20

7 Summary

The determined E_n value for the comparison measurement on pH is less than 0.3.

The comparison measurement was passed successfully by the participant SNM INDECOPI.

Bitterfeld-Wolfen, 29th April 2011

Diana Jehnert

Deputy Head of DKD-K-06901

Dr. Barbara Werner

Head of DKD-K-06901